

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

DC/DC Converter

Name and address of the applicant

MINMAX TECHNOLOGY CO LTD
18 SIN-SIN RD, AN-PING INDUSTRIAL DISTRICT,
TAINAN CITY, 702 TAIWAN

Name and address of the manufacturer

MINMAX TECHNOLOGY CO LTD
18 SIN-SIN RD, AN-PING INDUSTRIAL DISTRICT,
TAINAN CITY, 702 TAIWAN

Name and address of the factory

Note: When more than one factory, please report on page 2

MINMAX TECHNOLOGY CO LTD
18 SIN-SIN RD, AN-PING INDUSTRIAL DISTRICT,
TAINAN CITY, 702 TAIWAN Additional Information on page 2

Ratings and principal characteristics

See Page 2



Trademark (if any)

Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

MAU01-xSyM
See Page 2

Additional information (if necessary may also be reported on page 2)

Risk Management was not included in this investigation.

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate

170400502 issued on 2017-08-22

This CB Test Certificate is issued by the National Certification Body



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA



UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK



UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN



UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-09-01

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-66385-UL

Model Details:
MAU01-xSyM (x can be 05, 12, 24, y can be 05, 12 or 15)

Ratings:

Input voltage	Input current	Model	Output voltage	Output current	Transformer P/N
4.5-5.5 Vdc	0.32-0.27A	MAU01-05S05M	5 Vdc	200 mA	0B-MAU01-05S05M-T1
		MAU01-05S12M	12 Vdc	84 mA	0B-MAU01-05S12M-T1
		MAU01-05S15M	15 Vdc	68 mA	0B-MAU01-05S15M-T1
10.8-13.2 Vdc	0.14-0.11A	MAU01-12S05M	5 Vdc	200 mA	0B-MAU01-12S05M-T1
		MAU01-12S12M	12 Vdc	84 mA	0B-MAU01-12S12M-T1
		MAU01-12S15M	15 Vdc	68 mA	0B-MAU01-12S15M-T1
21.6-26.4 Vdc	0.07-0.06A	MAU01-24S05M	5 Vdc	200 mA	0B-MAU01-24S05M-T1
		MAU01-24S12M	12 Vdc	84 mA	0B-MAU01-24S12M-T1
		MAU01-24S15M	15 Vdc	68 mA	0B-MAU01-24S15M-T1

Additional Information:
Additionally evaluated to EN 60601-1:2006 / A1:2013 / A12:2014.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/names

Date: 2017-09-01

Signature: 
Jan-Erik Storgaard