



Ref. Certif. No.

SI-5265IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**Product
Produit

DC/DC Converter

Name and address of the applicant
Nom et adresse du demandeurMinmax Technology Co. Ltd.
18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702
TaiwanName and address of the manufacturer
Nom et adresse du fabricantMinmax Technology Co. Ltd.
18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702
TaiwanName and address of the factory
Nom et adresse de l'usineMinmax Technology Co. Ltd.
18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702
Taiwan*Note: When more than one factory, please report on page 2*
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page Additional Information on page 2Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

See page 2 for details.

Trademark (if any)
Marque de fabrique (si elle existe)

MINMAX

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

/

Model / Type Ref.
Ref. De typeMDHU100 series
See page 2 for details.Additional information (if necessary may also be reported
on page 2)
Les informations complémentaires (si nécessaire, peuvent
être indiqués sur la 2^{ème} pageThe risk management requirements of the standard were not
addressed. Additional Information on page 2A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à laIEC 60601-1:2005 (3rd Ed.) + A1:2012As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

T223-0216/16, dated 2016-03-16

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**SIQ Ljubljana
Tržaška cesta 2, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si
Product Certification Body is accredited by Slovenian Accreditation, Reg. No.: CP-001

Date: 2016-03-16

Signature: Bojan Pečavar

Ratings and principal characteristics:

Model name	Input voltage range	Rated input current	Output voltage	Output current
MDHU102	Nominal: 5 Vdc	600 mA	5 Vdc	400 mA
MDHU104	Range: 4,5-5,5 Vdc	600 mA	12 Vdc	165 mA
MDHU105	Input fuse: 1000 mA	600 mA	15 Vdc	133 mA
MDHU108		600 mA	± 12 Vdc	± 83 mA
MDHU109		600 mA	± 15 Vdc	± 66 mA
MDHU112		Nominal: 12 Vdc	250 mA	5 Vdc
MDHU114	Range: 10,8-13,2 Vdc	250 mA	12 Vdc	165 mA
MDHU115	Input fuse: 500 mA	250 mA	15 Vdc	133 mA
MDHU118		250 mA	± 12 Vdc	± 83 mA
MDHU119		250 mA	± 15 Vdc	± 66 mA
MDHU122		Nominal: 24 Vdc	135 mA	5 Vdc
MDHU124	Range: 21,6-26,4 Vdc	135 mA	12 Vdc	165 mA
MDHU125	Input fuse: 200 mA	135 mA	15 Vdc	133 mA
MDHU128		135 mA	± 12 Vdc	± 83 mA
MDHU129		135 mA	± 15 Vdc	± 66 mA

Additional information (if necessary)**Information complémentaire (si nécessaire)**

This CB test certificate substitutes previously issued CB test certificate No. SI-4536 dated 2015-06-10 due to test report update.

Date: 2016-03-16

Signature: Bojan Pečavar

