

## DK-84099-UL IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) **CB SCHEME CB TEST CERTIFICATE** Product DC/DC Converter Name and address of the applicant MINMAX TECHNOLOGY CO LTD 18 SIN-SIN RD, AN-PING INDUSTRIAL DISTRICT, TAINAN CITY 702, TAIWAN Name and address of the manufacturer MINMAX TECHNOLOGY CO LTD 18 SIN-SIN RD, AN-PING INDUSTRIAL DISTRICT, TAINAN CITY 702, TAIWAN Name and address of the factory MINMAX TECHNOLOGY CO LTD 18 SIN-SIN RD, AN-PING INDUSTRIAL DISTRICT, TAINAN Note: When more than one factory, please report on page 2 CITY 702, TAIWAN Additional Information on page 2 Ratings and principal characteristics Input: 9-18 Vdc, 1.41-0.7 A for model MIW10-12S033M, MIW10-12S05M, MIW10-12S051M, MIW10-12S12M, MIW10-12S15M, MIW10-12S24M, MIW10-12D12M, MIW10-12D15M Trademark (if any) See test report for details. <>>MINMAX<sup>®</sup> ₀ MINMAX<sup>®</sup> Type of Customer's Testing Facility (CTF) Stage used Model / Type Ref. MIW10-xDzM, MIW10-xSyM See Page 2 Additional information (if necessary may also be Additional Information on page 2 reported on page 2) A sample of the product was tested and found IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005 to be in conformity with

As shown in the Test Report Ref. No. which forms part of this Certificate

This CB Test Certificate is issued by the National Certification Body



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

181200602 issued on 2019-05-28

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-05-30

Signature:

 $\boxtimes$ 

Jan-Erik Storgaard

|--|

Ref. Certif. No.

## DK-84099-UL

## Model Details:

MIW10-xDzM,MIW10-xSyM (x can be 12, 24 or 48, y can be 033, 05, 051, 12, 15 or 24, z can be 12 or 15)

Additional Information: Risk Management was not included in this investigation. National Difference specified in the CB Test Report.

## Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-05-30

Signature:

Χ

Jan-Erik Storgaard

for but Superial